

Instructions for Wearers

AFTER YOUR

PARAGON CRT[®]

OR

PARAGON CRT[®] 100

CONTACT LENSES

FOR

CONTACT LENS CORNEAL REFRACTIVE THERAPY

HAVE BEEN FITTED



Just watch us[®]

Instructions for Wearers of
Paragon CRT[®] (paflucocon B)
or
Paragon CRT[®] 100 (paflucocon D)

Contact Lenses for Contact Lens Corneal Refractive Therapy

Patient Name: _____

Prescribed Lens: _____

Dr. _____

Address _____

Phone _____

CAUTIONS: Federal (US) law restricts this device to sale by, or on the order of a licensed practitioner.

Contact lenses for Corneal Refractive Therapy should be fitted only by a contact lens fitter trained and certified in the fitting of conventional (non-reverse geometry), sigmoid geometry, and reverse geometry contact lenses. Nonsterile. Clean and condition lenses prior to use.

The lens is shipped dry; or, wet shipped in solution. This solution contains poloxamine, hydroxyalkylphosphonate, boric acid, sodium borate, sodium chloride, hydroxypropylmethyl cellulose, Glucam and preserved with chlorhexidine gluconate (0.003%), polyaminopropyl biguanide (0.0005%). The case, packing slip or invoice is marked with the central base curve radius, diameter, dioptric power, overall diameter, Return Zone Depth, Landing Zone Angle, center thickness, serial number, ship date and color of the lens.

TABLE OF CONTENTS

	<u>Page</u>
Precautions	1
Adverse Effects (Problems And What To Do)	3
Personal Cleanliness And Lens Handling	3
Preparing The Lens For Wearing	3
Handling The Lens	4
Placing The Lens On The Eye	4
Centering The Lens	4
Removing The Lens	5
Caring For Your Lenses	5
Basic Instructions	5
Lens Deposits And Use Of Enzymatic Cleaning Procedures	7
Care For A Sticking (Nonmoving) Lens	7
Emergencies	7
Wearing Schedule	7

PRECAUTIONS

General

Clinical studies have demonstrated that Paragon CRT® and Paragon CRT® 100 contact lenses manufactured from Paragon HDS® and Paragon HDS® 100 respectively are safe and effective for their intended use. However, due to the small number of patients enrolled in the clinical investigation of lenses, all refractive powers, design configurations, and lens parameters available in the lens materials were not evaluated in significant numbers. This is especially true for adolescent subjects in this investigation. Consequently, when selecting an appropriate lens design and parameters, the eye care practitioner should consider all characteristics of the lens that can affect lens performance and your ocular health; including, oxygen permeability, wettability, central and peripheral thickness, and optic zone diameter.

The potential impact of these factors on your ocular health must be carefully weighed against your need for refractive reduction; therefore, your continuing ocular health and lens performance on your eye should be carefully monitored by your eye care practitioner. Corneal edema is more prevalent when the lens is used in high altitudes.

Each Paragon CRT® and Paragon CRT® 100 lens is supplied nonsterile in an individual plastic case. The lens is shipped dry; or, wet shipped in solution. This solution contains poloxamine, hydroxyalkylphosphonate, boric acid, sodium borate, sodium chloride, hydroxypropylmethyl cellulose, Glucam and preserved with chlorhexidine gluconate (0.003%) and polyaminopropyl biguanide (0.0005%). If you have experienced a prior history of allergy to any of these ingredients, remove the lens from the solution and soak the lens 24 hours in sterile unpreserved saline prior to cleaning, disinfecting and dispensing.

Never reuse the solution. You may store the lens in the unopened container until ready to dispense, up to a maximum of twenty-five (25) days from the Ship Date (see Packing Slip). If the lens is stored for longer periods of time, it should be cleaned and disinfected with an FDA approved product. Follow the directions on the selected disinfecting solution regarding prolonged storage.

Patient

Solution Precautions

- Different solutions cannot always be used together and not all solutions are safe for use with all lenses. Use only recommended solutions with your Paragon CRT® and Paragon CRT® 100 contact lenses.
- Do not heat the wetting/soaking solution and lenses.
- Always use fresh unexpired lens care solutions.
- Always follow directions in the package inserts of the contact lens solutions used.
- Use only a chemical lens care system. Use of a heat (thermal) lens care system can cause damage by warping your contact lenses.
- Sterile unpreserved solutions, when used, should be discarded after the time specified in the labeling directions.
- Do not use saliva, tap water or anything other than the recommended solutions for lubricating or wetting lenses.
- Always keep the lenses completely immersed in the recommended storage solution when the lenses are not being worn (stored).

Handling Precautions

- Always wash and rinse hands before handling lenses. Do not get cosmetics, lotions, soaps, creams, deodorants, or sprays in the eyes or on the lenses. It is best to put on lenses before putting on makeup. Water-base cosmetics are less likely to damage lenses than oil-base products.
- Be certain that your fingers and hands are free of foreign material before touching your contact lenses, as microscopic scratches of the lenses may occur causing distorted vision and/or injury to the eye.
- Carefully follow the handling, insertion, removal, cleaning, disinfecting, storing and wearing instructions in this booklet and those prescribed by your eye care practitioner.
- Always handle your lenses carefully and avoid dropping them.
- Never use tweezers or other tools to remove your lenses from the lens container unless specifically indicated for that use. Pour your lens into your hand.
- Do not touch the lens with your fingernails.
- To minimize lens warpage during cleaning, the lenses should be cleaned in the palm of the hand rather than between the thumb and fingers.

Lens Wearing Precautions

- CAUTION: Nonsterile. Clean and condition lenses prior to use.
- If the lens sticks (stops moving) on the eye, follow the recommended directions on “Care For A Sticking (Nonmoving) Lens” in this patient information booklet. The lens should move freely on the eye for the continued health of the eye. If nonmovement of the lens continues, you should immediately consult your eye care practitioner or attending hospital emergency room physician.
- Never wear your contact lenses beyond the period recommended by your eye care practitioner.
- Avoid, if possible, all harmful or irritating vapors and fumes when wearing lenses.
- If aerosol products such as sprays are used while wearing lenses, exercise caution and keep eyes closed until the spray has settled.

Lens Case Precautions

- Contact lens cases can be a source of bacterial growth. To prevent contamination and to help avoid serious eye injury, always empty and rinse the lens case with fresh, sterile rinsing solution and allow to air dry.
- Lens cases should be replaced at regular intervals as recommended by the lens case manufacturer or eye care practitioner.

Topics To Discuss With The Eye Care Practitioner

- Ask your eye care practitioner about wearing your lenses during sporting activities.
- Always contact your eye care practitioner before using any medicine in your eyes.

- As with any contact lens, follow-up visits are necessary to assure the continuing health of your eyes. You should be instructed as to a recommended follow-up schedule.

Who Should Know That You Are Wearing Contact Lenses

- Inform your doctor (health care practitioner) about being a contact lens wearer.
- If you choose to wear your lenses while at work always inform your employer of being a contact lens wearer. Some jobs may require the use of eye protection equipment or may require that you not wear contact lenses during work hours.

ADVERSE EFFECTS (PROBLEMS AND WHAT TO DO)

You should be informed that the following problems may occur.

- Eyes stinging, burning, itching (irritation), or other eye pain
- Comfort is less than when lens was first placed on eye
- Feeling of something in the eye such as a foreign body or scratched area
- Excessive watering (tearing) of the eyes
- Unusual eye secretions
- Redness of the eyes
- Reduced sharpness of vision (poor visual acuity)
- Blurred vision, rainbows, or halos around objects
- Sensitivity to light (photophobia)
- Dry eyes

If you notice any of these conditions: **IMMEDIATELY REMOVE YOUR LENSES.**

If the discomfort or problem stops, then look closely at the lens. If the lens is in any way damaged, **DO NOT** put the lens back on your eye. Place the lens in the storage case and contact your eye care practitioner. If the lens has dirt, an eyelash, or other foreign objects on it, or the problem stops and the lens appears undamaged, you should thoroughly clean, rinse and disinfect the lens; then reinsert it. If the problem continues, you should **IMMEDIATELY** remove the contact lenses and consult your eye care practitioner.

When any of the above problems occurs, a serious condition such as infection, corneal ulcer, neovascularization, iritis, persistent stromal edema or GPC (giant papillary conjunctivitis) may be present. You should keep the lens off the eye and seek immediate professional identification of the problem and prompt treatment to avoid serious eye damage, including corneal scarring, opacification, blindness or loss of eye.

PERSONAL CLEANLINESS AND LENS HANDLING

Preparing The Lens For Wearing

It is essential that you learn and use good hygienic methods in the care and handling of your new lenses. Cleanliness is the first and most important aspect of proper contact lens care. In particular, your hands should be clean and free of any foreign substance when you handle your lenses. The procedures are:

- Always wash your hands thoroughly with a mild soap, rinse completely, and dry with a lint-free towel before touching your lenses.
- Avoid the use of soaps containing cold cream, lotion, or oily cosmetics before handling your lenses, since these substances may come into contact with the lenses and interfere with successful wearing.
- To avoid damaging your lenses, handle them with your fingertips, and be careful to avoid contact with your fingernails. It is helpful to keep your fingernails short and smooth.

- Start correctly by getting into the habit of always using proper hygienic procedures so that they become automatic.

Handling The Lens

Develop the habit of always working with the same lens first to avoid mix-ups.

Remove the lens from its storage case and examine it to be sure that it is moist, clean, and free of any nicks and cracks.

Placing The Lens On The Eye

Work over a table, upon which is placed a clean towel. Do not place lenses on the eye while working over a sink. For the right eye:

- Wet your right index finger with a drop of conditioning solution and place the contact lens front side down on your right index finger.
- Place the second finger of the left hand on the middle of the upper lid and press firmly upward.
- Place the second finger of the right hand on the lower lid and press firmly downward.
- Stare into a mirror as though looking through the second finger holding the contact lens. You will later learn to do this without a mirror.
- Slowly move the hand to advance the forefinger with the contact lens towards the cornea until the lens touches the cornea and release the lids.
- Release the lid and close the eye for a few seconds.

Repeat procedure for the left eye.

There are other methods of lens placement. If this method is difficult for you, your eye care practitioner will provide you with an alternate method.

Note: If after placement of the lens your vision is blurred, check for the following:

1. The lens is not centered on the eye (see “Centering The Lens”, next section in this booklet).
2. If the lens is centered, remove the lens (see “Removing The Lens” section) and check for the following:
 - a. Cosmetics or oils on the lens. Clean, rinse, disinfect, and place on the eye again.
 - b. The lens is on the wrong eye.

If you find that your vision is still blurred after checking the above possibilities, remove both lenses and consult your eye care practitioner.

Centering The Lens

Very rarely, a lens that is on the cornea will be displaced onto the white part of the eye during lens wear. This may also occur during placement and removal of the lenses if the correct techniques are not performed properly. To center a lens follow this procedure.

- First locate the lens by pulling away the lids.

- After the lens is found, gently press on the lid over the lens while looking away from the direction of the lens.
- Next look back towards the lens.

Removing The Lens

Always remove the same lens first.

- Wash, rinse, and dry your hands thoroughly.
- Work over a table with a clean towel. Do not remove lenses over a sink.
- Place the right index finger at the outer corner of the eye.
- Place the left hand cupped below the eye.
- Open the eyes wide as if to stare.
- Continue to keep the eyes open and pull the lids sideways away from nose.
- Blink quickly and firmly.

Remove the second lens by following the same procedure.

OPTIONAL METHOD: If you choose to use a DMV Classic lens remover device, compress the handle of the DMV between your thumb and index finger while looking straight ahead carefully center it squarely on the contact lens. When the Classic has been seated on the lens, release the compression pressure and gently tip the remover down to retract the lens. Always be aware of the position of the lens on the eye before applying the Classic to the lens. Then, squeeze the stem of the remover again to release the suction and slide the lens from the remover between your thumb and forefinger. Remove second lens following the same procedure. Follow the manufacturer's instructions for cleaning your DMV Classic. For all other lens removing devices follow manufacturer's instructions.

Follow the required lens care procedures described under the heading: CARING FOR YOUR LENSES.

Note: If these methods of removing your lens are difficult for you, your eye care practitioner will provide you with an alternate method.

CARING FOR YOUR LENSES

Basic Instructions

For continued safe and comfortable wearing of your lenses, it is important that you clean and rinse, then disinfect your lenses after each removal using the care regimen recommended by your eye care practitioner. Cleaning and rinsing are necessary to remove mucus, secretions, films, or deposits, which may have accumulated during wearing. The ideal time to clean, rinse and disinfect your lenses is immediately after wearing them. Disinfecting is necessary to destroy harmful germs.

You should adhere to a recommended care regimen. Failure to follow the regimen may result in development of serious ocular complications as discussed in the WARNINGS section of the Package Insert.

When you first receive your lenses, practice how to put the lenses on and how to remove them while you are in your eye care practitioner's office. At that time you will be provided with a recommended cleaning and disinfection regimen and, instructions and warnings for lens care, handling, cleaning, and disinfection. Your eye care practitioner should instruct you about appropriate and adequate procedures and products for your use.

For safe contact lens wear you should know and always practice your lens care routine.

- Always wash, rinse, and dry hands before handling contact lenses.
- Always use fresh unexpired lens care solutions.
- Use the recommended system of lens care, which is chemical (not heat) and carefully follow instructions on solution labeling. Different solutions cannot always be used together and not all solutions are safe for use with all lenses. **Do not alternate or mix lens care systems unless indicated on solution labeling.**
- Always remove, clean, rinse, enzyme and disinfect your lenses according to the schedule prescribed by your eye care practitioner. The use of an enzyme or any cleaning solution does not substitute for disinfection.
- To avoid contamination, do not use saliva, tap water or anything other than the recommended solutions for lubricating or rewetting your lenses. Do not put lenses in your mouth.

Your eye care practitioner will recommend his/her preferred, FDA approved lens care solutions for the cleaning, disinfection, storage and lubrication of your Paragon CRT® and Paragon CRT® 100 contact lenses.

Follow the instructions provided with each lens care solution. Failure to adhere to these procedures may result in the development of serious ocular complications. A patient should not switch from one care system to another unless it has been determined by the eye care practitioner that this is necessary. Do not mix or alternate the disinfection and storage systems unless so indicated on the product label.

Always wash and rinse your hands thoroughly before handling your contact lenses.

1. Clean

Clean one lens first (always start with the same lens first to avoid mix-ups). Place the lens, front side down, in the palm of the hand and apply several drops of cleaning solution. Using the index finger of the other hand, apply slight pressure in a swirling motion for the time recommended by the cleaning solution manufacturer. Do not clean the lens by rubbing it between the thumb and index fingers, as this may cause lens warpage.

2. Rinse

Rinse the lens thoroughly as recommended by your lens care product manufacturer to remove the cleaning solution, mucus, and film from the lens surface. Place that lens into the correct chamber of the lens storage case. Then repeat the clean and rinse procedure for the second lens.

3. Disinfect

After cleaning and rinsing the lenses disinfect them by using the system recommended by your eye care practitioner and/or the lens manufacturer. Follow the instructions provided in the disinfection solution labeling.

4. Storage

To store lenses, disinfect and leave them in the closed case until ready to wear. If lenses are not to be used immediately following disinfection, you should consult the storage solution package insert or your eye care practitioner for information on storage of your lenses.

Always keep your lenses completely immersed in a recommended disinfecting/conditioning solution when the lenses are not being worn. If you discontinue wearing your lenses, but plan to begin wearing them again after a few weeks, ask your eye care practitioner for a recommendation on how to store your lenses.

Note: Paragon CRT® and Paragon CRT® 100 Contact Lenses for Corneal Refractive Therapy cannot be heat (thermally) disinfected.

5. Care of Your Lens Case

Contact lens cases can be a source of bacteria growth. After removing your lenses from the lens case, empty and rinse the lens storage case with solution(s) recommended by the lens case manufacturer; then allow the lens case to air dry. When the case is used again, refill it with fresh disinfecting solution. Lens cases should be replaced at regular intervals as recommended by the lens case manufacturer or your eye care practitioner.

6. Lubricating/Rewetting

Your eye care practitioner will recommend a lubricating/rewetting solution. Lubricating/Rewetting solutions can be used to rewet (lubricate) your lenses while you are wearing them to make them more comfortable.

Lens Deposits And Use Of Enzymatic Cleaning Procedure

Your eye care practitioner may recommend enzyme cleaning. Enzyme cleaning removes protein deposits on the lens. These deposits cannot be removed with regular cleaners. Removing protein deposits is important for the well-being of your lenses and eyes. If these deposits are not removed, they can damage the lenses and cause irritation.

Enzyme cleaning does not replace routine cleaning and disinfecting. For enzyme cleaning, you should carefully follow the instructions in the enzymatic cleaning labeling.

Care For A Sticking (Nonmoving) Lens

If the lens sticks (stops moving) or cannot be removed, you should apply 2 drops of the recommended lubricating or rewetting solution directly to the eye and wait until the lens begins to move freely on the eye before removing it. If nonmovement of the lens continues after 30 minutes, you should IMMEDIATELY consult your eye care practitioner or attending hospital emergency room physician.

EMERGENCIES

If chemicals of any kind (household products, gardening solutions, laboratory chemicals, etc.) are splashed into your eyes, you should:

- FLUSH YOUR EYES IMMEDIATELY WITH TAP WATER.
- REMOVE YOUR LENSES.
- IMMEDIATELY CONTACT YOUR EYE CARE PRACTITIONER OR VISIT A HOSPITAL EMERGENCY ROOM WITHOUT DELAY.

WEARING SCHEDULE

Typically, your practitioner will start your overnight wear the first night, or your practitioner may initiate your lens wear on a daytime schedule. If your practitioner places you on an overnight wear schedule, you should place the lens in your eye 15 to 20 minutes before going to sleep. A well fit lens provides for centration with the closed eye. The effects of lid interaction on blinking and gravity may result in lens decentration during open eye wear.

Be aware, “when in doubt, take it out”. It is important that the new wearer not sleep in a lens that has a significant foreign body sensation. In the event of foreign body sensation, remove the lens, clean and re-wet it; and, again place the lens in your eye. If the sensation continues, remove the lens. The lens should not be worn.

Your practitioner will schedule a follow-up evaluation the morning after the first overnight wear. The visit is best scheduled within a few hours of awakening and you should report with your lenses in place. This visit provides an excellent opportunity to evaluate lens centration and potential lens adherence.

Assuming the absence of clinical signs and complications, you will be instructed to continue overnight wear of the lenses until the next scheduled follow-up visit.

The cornea normally changes within five to eight hours of wear. Your practitioner should modulated your wearing time to determine the MINIMUM wear required for myopic reduction. The average wearing time is between 8 and 10 hours. Attempt to maintain wearing time at this minimum level.

Myopic Reduction Maintenance Lens (Retainer Lens) Schedule

With the Paragon CRT® and Paragon CRT® 100 contact lenses, the lens used to achieve refractive therapy is usually the lens used to maintain achieved correction.

Note: To maintain the Contact Lens Corneal Refractive Therapy effect of myopia reduction overnight lens wear must be continued on a prescribed schedule. Failure to do so can affect daily activities (e.g., night driving), visual fluctuations and changes in intended correction.

Manufacturer:

Paragon Vision Sciences, Inc.
947 E. Impala Avenue
Mesa, Arizona 85204-6619

1-800-528-8279
1-480-892-7602
1-480-926-7369 FAX